



**CIPET : Institute of Plastics Technology (IPT), Ahmedabad**

Department of Chemicals & Petrochemicals, Ministry of Chemicals & Fertilizers

Plot No. 630, Phase - IV, GIDC, VATVA, Ahmedabda - 382 445, GUJARAT

E-mail: cipetahmd@gmail.com, ahmedabad@cipet.gov.in



**TECHNICAL BID**

**Name of Contractor/Bidder**

e-TENDER NO **CIPET/IPT/AHMD/ADMN/BDL/18-19/(05)**

ITEM DESCRIPTION **TESTING EQUIPMENTS FOR ESTABLISHMENT OF BIO-DEGRADABLE LABORATORY :  
Incubator type Biodegradation Set-up with necessary Glassware**

Sl. No.	PARTICULARS	REMARKS
I	Name of the Supplier / Manufacturer	
II	Postal address	
III	Telephone No. with STD code	
IV	Fax with STD code	
V	Name of Contact person	
VI	Mobile No.	
VII	e-mail ID	
	<b>BELOW GIVEN DOCUMENTS TO BE SCANNED AND UPLOADED IN THE WEBSITE <a href="http://www.tenderwizard.com/CIPET">www.tenderwizard.com/CIPET</a> WITHIN THE PERIOD OF SUBMISSION.</b>	"YES" OR "NO"
1	<b>Certificates:</b> i) Registration Certificate if Any, with DPS/DGS&D/NSIC/MSME ii) PAN (Permanent Account Number) Registration iii) Certificates of Registration for Sales Tax/ GST/VAT/ WCT and Service Tax iv) Authorization certificate from principal (if not manufacturer)	
	<b>Commercial Terms for Quoted items (Please Provide Commercial terms and conditions in the below form)</b>	
2	Validity of offer	
3	Price Shall be firm throughout out the currency of contract, in the event of placement of purchase order.	
4	Terms of Payment:- i) 80% on submission of the proof of receipt of the machine/equipment from CIPET Centre along with original invoice. ii) 20% on submission of the installation report obtained from the respective centre heads.	
5	<b>Price Term (Free Delivery/ Safe Door Delivery)</b>	
6	Delivery period (in days)	
7	Guaranty / Warranty(Please Indicate the Period)	
	<b>QUESTIONNAIRE TO BE FILLED BY BIDDER IN AND SENT ALONG WITH OFFER DULY SIGNED</b>	
8	Whether All Documents Relating to tender Viewed	

**Technical Specification**

**1) Incubator type Biodegradation Set-up with necessary Glassware - 2 NOS.**

		Remarks (Yes/No)
	<b>MAKE &amp; MODEL</b> alongwith Ref. No. of Catalogue enclosed	
Auto-titrator titrations in for performing all types aqueous and non-aqueous of potentiometric phases.		
1	<b>Make &amp; Model</b>	Standard
2	Temperature Controller	Digital
3	Temperature Range	Ambient to 80 Deg. C.
4	<u>Temperature accuracy</u>	+/- 1 Deg. C.
	An integrally attached control panel accommodating digital temperature controller, safety thermostat, indicating lamps to be provided.	
	<b>Type</b>	<b>Incubator Type</b>
	Flow Meter for Incubator	24 Nos

	Composting Vessel of 3 litre and 5 litre capacity with air tight fitting.	12 Nos each	
	Cylindrical shape SS mesh type filter	12 Nos	
	Calcium Hydroxide filter column	1 No	
	Multi Storage Rack with Castor Wheels & with provision for flow meter attachment	1 No	
	Atleast 5000 ml Bottle with CorkFittings	36 Nos	
	Silicon Hose	200 Meters	
	Oil Free Compressor 2 HP	1 No.	
	Apparatus should confirms to	ASTM D 5338	