

Frequently Asked Questions (FAQ3)**1. Name of the Equipment: CIPET_HO/Pur/Ten.03/2020-21/Rockwell Hardness Tester**

S. No.	Query	CIPET Response
1	In Rockwell Hardness tester, you've given the scale, but you've to mention the test blocks needed. In standard delivery HRA, HRB, HRC each 2 blocks will be given. There are so many ranges are available, which range you've required to be given in the specification. If we quote for all the ranges then the price will be very high. Otherwise the test block prices may be quoted in Optional item . If these details are not given , you cannot compare the bids	Standard delivery blocks of HRA, HRB, HRC each 2 blocks with Calibration certificates traceable to NPL/NIST may be supplied

2. Name of the Equipment: CIPET_HO/Pur/Ten.03/2020-21/24/ Automized Izod / Charpy Impact Tester with Notch Cutter as per ASTM

S. No.	Query	CIPET Response
1	In the Impact Testing Machine, capacity of the Izod and Charpy weights to be mentioned. There are so many different hammers available for the same. If we include all, then the price of the Unit will be very expensive. Kindly confirm the weights required.	In the bid pendulum capacity is described as below: Aerodynamic compound pendulum with facility to increase/decrease pendulum capacity with addition/removal of weights enabling user to change to desired capacity quickly & easily. Basic pendulum capacity: 2.82J, with addition of weights, user should be able to change it to 25J (Standard)