

CORRIGENDUM

Tender No.: CIPET_HO/Pur./TEN.02/2019-20

1) Tool Room Machinery

Sl.No. 3 CNC Lathe (CNC Turning Centre)

7.1 Read as - Siemens / Fanuc / Equivalent (Latest with complete module)

2) Testing Equipment

Sl. No. 20 Deep Freezer

Internal Volume - 185 Ltrs

Internal Dimensions W x D x H (mm) - 510 x 510 x 730

3) Testing Equipment

Sl. No. 6 & 7 Universal Testing Machine 50 kN & 25 kN

1) What is percentage of elongation of sample ?

The machine is supposed to carryout testing on various type of Plastic specimens and Product samples as per IS and Internal Standards as mentioned in the tender specification. The sample involves FRP sheets, Plastic pipe cut pieces moulded specimens of commodity, Engineering Plastics as per different standards.

2) What is max elongation in mm ?

As mentioned in cross head travel requirement.

3) What is the gauge length of sample ?

Gauge length varies (from 25 mm to 250 mm) depending on the test method.


Principal Director (T)